	Application No.	Applicant(s)	
Nation of Allowshills.	10/038,563	RIPICH ET AL.	
Notice of Allowability	Examiner	Art Unit	
	Jessica R. Baxter	3731	
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.			
1. This communication is responsive to <u>Pre-Appeal Brief Conference Request of 8/18/2005</u> .			
2.  The allowed claim(s) is/are <u>17-20 and 22-31</u> .			
3.			
Attachment(s)  1. Notice of References Cited (PTO-892)  2. Notice of Draftperson's Patent Drawing Review (PTO-948)  3. Information Disclosure Statements (PTO-1449 or PTO/SB/O Paper No./Mail Date  4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	8. Examiner's Statements 9. Other EDUARDO	(PTO-413), te ment/Comment ent of Reasons for Allo	

Application/Control Number: 10/038,563 Page 2

## **REASONS FOR ALLOWANCE**

- 1. The following is an examiner's statement of reasons for allowance:
- 2. The prior art fails to disclose or suggest a tongue cleaning device comprising a head having a scraping wall with a front surface facing the handle of the body, the front surface defining a debris retention recess; the debris retention recess being defined in the front surface of the scraping wall. This limitation, in combination with all of the limitations of claims 17, 23 or 28 makes the claims allowable over the prior art.
- 3. As pointed out by the applicant in the response (Pre-Appeal Brief Conference Request), filed August 18, 2005, the debris retention recess of Ripich et al. '228 is not defined in the front surface of the scraping wall. The front surface of the device of Ripich is defined by the placement of two walls (41 and 31) and not by a recess entirely defined in the front surface of the wall. These two walls of Ripich do not meet the limitations claimed by applicant; wherein the recess is "defined in the scraping wall" or "in the front surface of the scraping wall". The front surfaces of walls (41 and 31) are smooth and there are no recesses or depressions defined in these surfaces. Therefore, Ripich et al. '228 does not anticipate the claimed invention.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R. Baxter whose telephone number is 571-272-4691. The examiner can normally be reached on M-F 8:30AM - 5:00PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Jessica R Baxter Examiner Art Unit 3731

) jrb

> EDUARDO C. ROBERT PRIMARY EXAMINER